


## Analysis of Tocopherol



Rs=6.3

Column	: Inertsil ODS-3V 5µm 150 × 4.6mm I.D.	Solutes	: 1. DL- $\alpha$ - Tocopherol(1.0mg/mL) 2. DL- $\alpha$ - Tocopherol acetate(1.0mg/mL)
Column Cat. No.	:		
Eluent	: CH <sub>3</sub> OH/H <sub>2</sub> O=98/2		
Flow rate	: 1.6 mL/min.		
Col. Temp.	: 30°C		
Detection	: UV 284 nm		
Injection Vol.	: 20µL		
Sample	: 1. DL- $\alpha$ - Tocopherol(1.0mg/mL) 2. DL- $\alpha$ - Tocopherol acetate(1.0mg/mL)		