


InertSearch for LC Data No. LA456-0000

Analysis of Polystyrene Standard



Column : Inertsil SIL-150A

Solutes

 $5 \mu m$

250 x 4.6 mm I.D.

Column Cat. No.: 5020-01012

Eluent : THF

Flow rate · 1.0 r

Flow rate : 1.0 mL/min.

Col. Temp. : Ambient

Detection : UV 254 nm

Injection Vol. :5 µL

Sample .

- :1. Polystylene (0.2 mg/mL, Mp 1,020,000)
- 2. Polystylene (0.2 mg/mL, Mp 115,000)
- 3. Polystylene (0.2 mg/mL, Mp 9,200)
- 4. Polystylene (0.2 mg/mL, Mp 950)