


## Analysis of Glycyrrhetic acid

(Under the Condition of the Japanese Standards of Quasi-drug Ingredients,  
Monoammonium  $\alpha$ -Glycyrrhizinate)



### Conditions

System : GL7700 HPLC system

Column : InertSustain C18  
(5  $\mu$ m, 250 x 4.6 mm I.D.)

Column Cat. No. : 5020-07346

Eluent : A) CH<sub>3</sub>CN  
B) 15 mM CH<sub>3</sub>COONH<sub>4</sub> in H<sub>2</sub>O  
A/B = 55/45, v/v

Flow Rate : 1.0 mL/min

Col. Temp. : 40 °C

Detection : UV 250 nm (PD7752 PDA Detector)


Injection Vol. : 20  $\mu$ L

Sample : Standard

### Analyte:

1.  $\alpha$ -Glycyrrhetic acid  
2.  $\beta$ -Glycyrrhetic acid

40 mg/L  
20 mg/L



## Analysis of Glycyrrhetic acid

(Under the Condition of the Japanese Standards of Quasi-drug Ingredients,  
Monoammonium  $\alpha$ -Glycyrrhizinate)




Fig.1 Calibration curve of  $\alpha$ -Glycyrrhetic acid




Fig.2 Calibration curve of  $\beta$ -Glycyrrhetic acid