
InertSearch™ for LC

Inertsil ™ Applications

Analysis of Proteins

Data No. LB601-0699



Conditions

Column	:	MonoSelect C18 for HTS (20 x 2.1 mm I.D.)
Column Cat. No.	:	5020-10811
Eluent	:	A) 0.1% TFA in CH_3CN B) 0.1% TFA in H_2O A/B = 25/75 – 2 min – 50/50, v/v
Flow rate	:	0.3 mL/min
Col. Temp.	:	40 ℃
Detection	:	UV 280 nm
Injection Vol.	:	1 μ L
Sample	:	Standard

Analyte:

1. Ribonuclease A
2. Cytochrome C
3. Lysozyme
4. Conalbumin

* An LC system optimized for high-throughput analysis was used.